4056 Blakie Road, London, Ontario, Canada N6L 1P7
T: (519) 652-5833 | F: (519) 652-3795
technologies@oes-inc.com
www.OEStechnologies.com

ISO 9001 2008
FM64157

ForcePak Update Instructions

1. Extract (UNZIP) and copy updated software files to a USB memory stick.

    (ForcePakUpdate.exe **Must be on root of USB Drive, not in a folder)**

2. When software is running on unit, insert USB stick.

3.   Select the update and then Install, the unit will run the update automatically and then reboot after complete.

4056 Blakie Road, London, Ontario, Canada N6L 1P7
T: (519) 652-5833 | F: (519) 652-3795
technologies@oes-inc.com
www.OEStechnologies.com

ISO 9001 2008
FM64157

4. The unit will now be at the latest level of software. To update the unit with any new updates, repeat previous steps.

List the install version as latest installed ⟵